A Company	

## Application No. Applicant(s) POLITZE ET AL. 10/647,318 Notice of Allowability **Art Unit** Examiner Tu T. Nguyen 2886 -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308. 1. This communication is responsive to 12/06/2006. 2. The allowed claim(s) is/are <u>1-5,7-12 and 20-31</u>. 3. Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). b) Some\* c) None of the: a) 🛛 All 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. \_\_\_\_\_. 3. Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)). \* Certified copies not received: \_\_\_\_\_. Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application. THIS THREE-MONTH PERIOD IS NOT EXTENDABLE. 4. A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient. 5. CORRECTED DRAWINGS (as "replacement sheets") must be submitted. (a) including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached 1) hereto or 2) to Paper No./Mail Date \_\_\_\_\_. (b) including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_. Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d). 6. DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL. Attachment(s) 1. Notice of References Cited (PTO-892) 5. Notice of Informal Patent Application 6. Interview Summary (PTO-413), 2. Notice of Draftperson's Patent Drawing Review (PTO-948) Paper No./Mail Date \_\_\_\_\_. 7. Examiner's Amendment/Comment 3. M Information Disclosure Statements (PTO/SB/08), Paper No./Mail Date 12/06/2006 4. Examiner's Comment Regarding Requirement for Deposit 8. Examiner's Statement of Reasons for Allowance of Biological Material 9. Other \_\_\_\_.

Application/Control Number: 10/647,318

Art Unit: 2886

## **REASONS FOR ALLOWANCE**

The following is an examiner's statement of reasons for allowance:

As per claim 1, the prior arts of record, taken alone or in combination, fail to disclose or render obvious the steps of emitting pulsed radiation of a second wavelength, which is shorter than the first wavelength along a second radiation axis into the measuring volume; measuring radiation scattered on particles located in the measuring volume under a forward scattering angle of more than 90 degrees and under a backward scattering angle of less than 90 degrees, wherein forward scattered radiations and backward scattered radiations of the first and second wavelengths are measured separately from each other, wherein the scattered radiations of the first and second wavelengths are measured on opposite sides of the measuring volume on a same main axis, which is structurally arranged and functionally operated as claimed.

As per claim 20, the prior arts of record, taken alone or in combination, fail to disclose or render obvious the steps of emitting pulsed radiation of a second wavelength, which is shorter than the first wavelength along a second radiation axis into the measuring volume; measuring radiation scattered on particles located in the measuring volume under a forward scattering angle of more than 90 degrees and under a backward scattering angle of less than 90 degrees, wherein forward scattered radiations and backward scattered radiations of the first and second wavelengths are measured separately from each other, wherein the scattered radiations of the first and second wavelengths are emitted into the measuring volume from opposite sides along

Application/Control Number: 10/647,318

Art Unit: 2886

coinciding radiation axes, which is structurally arranged and functionally operated as claimed.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Tu T. Nguyen whose telephone number is (571) 272-2424. The examiner can normally be reached on T-F 7:30-5:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Tarifur Chowdhury can be reached on (571) 272-2800 Ext. 86. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 2886

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Tu T. Nguyen
Primary Examiner
Art Unit 2886

Durguyen

02/28/2007